

# FLUKE®

# 63/66/68

## Infrared Thermometers

**Users Manual**

PN 2149032

September 2004 Rev. 3, 6/09

© 2004-2009 Fluke Corporation. All rights reserved. Printed in China.

Specifications are subject to change without notice.

All product names are trademarks of their respective companies.

## **LIMITED WARRANTY AND LIMITATION OF LIABILITY**

This Fluke product will be free from defects in material and workmanship for two years from the date of purchase. This warranty does not cover fuses, disposable batteries, or damage from accident, neglect, misuse, alteration, contamination, or abnormal conditions of operation or handling. Resellers are not authorized to extend any other warranty on Fluke's behalf. To obtain service during the warranty period, contact your nearest Fluke authorized service center to obtain return authorization information, then send the product to that Service Center with a description of the problem.

**THIS WARRANTY IS YOUR ONLY REMEDY. NO OTHER WARRANTIES, SUCH AS FITNESS FOR A PARTICULAR PURPOSE, ARE EXPRESSED OR IMPLIED. FLUKE IS NOT LIABLE FOR ANY SPECIAL, INDIRECT, INCIDENTAL OR CONSEQUENTIAL DAMAGES OR LOSSES, ARISING FROM ANY CAUSE OR THEORY.** Since some states or countries do not allow the exclusion or limitation of an implied warranty or of incidental or consequential damages, this limitation of liability may not apply to you.

Fluke Corporation  
P.O. Box 9090  
Everett, WA 98206-9090  
U.S.A.

Fluke Europe B.V.  
P.O. Box 1186  
5602 BD Eindhoven  
The Netherlands

# ***Table of Contents***

| <b>Title</b>  | <b>Page</b> |
|---|-------------|
| Introduction.....   | 1           |
| Contacting Fluke .....  | 1           |
| Safety Information .....  | 2           |
| Symbols and Safety Markings.....                                  | 3           |
| Features .....  | 5           |
| Display .....   | 7           |
| Display (63).....   | 8           |
| Display (66/68).....  | 8           |
| Buttons (66/68).....  | 9           |
| How the Thermometers Work.....                                    | 9           |
| Operating the Thermometer .....                                   | 10          |
| Temperature Measurement.....                                      | 10          |
| Locating a Hot or Cold Spot.....                                  | 10          |
| Distance and Spot Size.....                                       | 11          |
| Field of View .....   | 13          |
| Emissivity .....  | 13          |
| Switching °C and °F .....   | 15          |
| Trigger Lock (66/68).....   | 15          |
| Backlight and Laser On/Off Switch (63) .....                      | 16          |
| Function Button Functions (66/68) .....                           | 16          |
| Selecting a Function (66/68) .....                                | 17          |
| Setting the High Alarm, Low Alarm, and<br>Emissivity (66/68)..... | 18          |

|  |    |
|--|----|
| Using a Contact Temperature Probe (PRB)<br>(66/68) ..... | 18 |
| HOLD .....   | 19 |
| Storing Data (66/68) .....                               | 19 |
| Recalling Data (66/68) .....                             | 19 |
| LOG Clear Function.....                                  | 19 |
| Maintenance .....  | 20 |
| Changing the Battery .....                               | 20 |
| Cleaning the Lens.....                                   | 20 |
| Cleaning the Housing .....                               | 20 |
| Troubleshooting .....                                    | 21 |
| Accessories.....   | 21 |
| CE Certification .....                                   | 22 |
| Specifications.....                                      | 22 |

## **Introduction**

The Fluke Models 63, 66, and 68 Infrared Thermometers (“the thermometers”) are for non-contact temperature measurement. These thermometers determine an object’s surface temperature by measuring the amount of infrared energy radiated by the object’s surface. See Figure 2.

## **Contacting Fluke**

To contact Fluke, call one of the following telephone numbers:

- Technical Support USA: 1-800-44-FLUKE (1-800-443-5853)
- Calibration/Repair USA: 1-888-99-FLUKE (1-888-993-5853)
- Canada: 1-800-36-FLUKE (1-800-363-5853)
- Europe: +31 402-675-200
- Japan: +81-3-3434-0181
- Singapore: +65-738-5655
- Anywhere in the world: +1-425-446-5500

Or, visit Fluke's website at [www.fluke.com](http://www.fluke.com).



To register your product, visit <http://register.fluke.com>.

To view, print, or download the latest manual supplement, visit <http://us.fluke.com/usen/support/manuals>.

## Safety Information

### Warning

A Warning identifies conditions and actions that pose hazards to the user. To avoid electrical shock or personal injury, follow these guidelines:

-  Do not point laser directly at eye or indirectly off reflective surfaces.
- Before using the thermometer inspect the case. Do not use the thermometer if it appears damaged. Look for cracks or missing plastic.
- Replace the batteries as soon as the battery indicator () appears.
- Do not use the thermometer if it operates abnormally. Protection may be impaired. When in doubt, have the thermometer serviced.
- Do not operate the thermometer around explosive gas, vapor, or dust.
- Do not connect the optional external probe to live electrical circuits.
- To avoid a burn hazard, remember that highly reflective objects will result in lower than actual temperature measurements.
- Do not use in a manner not specified by this manual or the protection supplied by the equipment may be impaired.

### Caution






To avoid damaging the thermometer or the equipment under test protect them from the following:

- EMF (electro-magnetic fields) from arc welders, induction heaters, etc.
- Static electricity
- Thermal shock (caused by large or abrupt ambient temperature changes - allow 30 minutes for thermometer to stabilize before use).
- Do not leave the thermometer on or near objects of high temperature.

## Symbols and Safety Markings

Table 1 and Figure 1 show various symbols and safety markings that are on the thermometers and in this manual.

**Table 1. Symbols**

| Symbol  | Explanation  |
|---|--|
|    | Risk of danger. Important information. See Manual.   |
|    | Hazardous voltage. Precedes warning  |
|    | Warning. Laser.  |
| CE  | Conforms to requirements of European Union and European Free Trade Association (EFTA)                        |
| °C  | Celsius  |
| °F  | Fahrenheit   |
|    | Battery  |
|  | Do not dispose of this product as unsorted municipal waste. Go to Fluke's website for recycling information. |




Figure 1. Safety Markings on the Thermometers



## **Features**

The thermometers include:

- Single-spot laser sighting
- MAX temperature display
- Backlit display
- Durable, ergonomic construction
- Tripod mount
- Hard case and wrist strap
- MAX, MIN, DIF, AVG temperature displays (66/68)
- Adjustable emissivity (66/68)
- High and low alarm (66/68)
- Data logging (66/68)
- Trigger Lock (66/68)
- Contact probe jack (66/68)





Figure 2. Infrared Thermometer

ame007.eps

## Display



ame001.eps

|   |  |
|---|--|
| ① | Backlight "On" symbol (63/66/68)   |
| ② | °C/°F symbol (Celsius/Fahrenheit) (63/66/68)   |
| ③ | High alarm and low alarm symbol (66/68)  |
| ④ | Maximum temperature display (63)<br>Temperature values for the MAX, MIN, DIF, AVG, HAL, LAL, PRB (66/68) |
| ⑤ | Icon for MAX (63/66/68)<br>Icon for MIN, DIF, AVG, HAL, LAL, PRB (66/68)                                 |
| ⑥ | LOG icon shows log mode for data storage (66/68)   |
| ⑦ | Live temperature value (63/66/68)  |
| ⑧ | SCAN or HOLD (63/66/68)  |
| ⑨ | Emissivity symbol and value (63/66/68)   |
| ⑩ | Low Battery and Laser "On" symbols (63/66/68)<br>Lock symbol (66/68)                                     |


**Display (63)**


In the SCAN mode, the backlit reading displays both the live temperature (⑦) and maximum temperature (④) in Celsius or Fahrenheit (②). The thermometer holds the last reading for 7 seconds after the trigger is released and “HOLD” appears (⑧).

**Display (66/68)**






In SCAN mode, both the live temperature (⑦) and the selected mode function (④,⑤) are displayed in Celsius or Fahrenheit (②). The thermometer holds the last reading for 7 seconds after the trigger is released and “HOLD” appears (⑧).

*Notes*

*For all units, when the battery is low,  appears but the thermometer will no longer function.*

*For the 66/68, to activate the laser and backlight, pull the trigger. Press  once to activate the backlight, twice to turn both laser and backlight on, and a third time to turn them off.*

## **Buttons (66/68)**

|   |   |
|---|---|
|  | Sets high and low alarm   |
|  | Up and down   |
|  | Cycles through the function loop  |
|  | Laser/Backlight on/off button (pull trigger and press button to activate laser/backlight) |
|  | LOG button (for storing data)   |

## **How the Thermometers Work**

Infrared thermometers measure the surface temperature of an opaque object. The thermometer's optics sense emitted, reflected, and transmitted energy, which is collected and focused onto a detector. The unit's electronics translate the information into a temperature reading which the unit displays. The laser is used for aiming purposes only. See Figure 3.




Figure 3. How the Thermometer Works


## **Operating the Thermometer**

### **Temperature Measurement**

To measure temperature, point the unit at an object and pull the trigger. Be sure to consider distance-to-spot size ratio and field of view. The laser is used for aiming only.

### **Locating a Hot or Cold Spot**

To find a hot or cold spot, aim the thermometer outside the desired area. Then, slowly scan across the area with an up and down motion until you locate the hot or cold spot. See Figure 4.



ame003.eps

**Figure 4. Locating a Hot or Cold Spot**

### ***Distance and Spot Size***

As the distance ( $D$ ) from the object being measured increases, the spot size ( $S$ ) of the area measured by the unit becomes larger. The relationship between distance and spot size for each unit is shown in Figure 5. The focal point of each unit is 914 mm (36 in). The spot sizes indicate 90 % encircled energy. See Figure 5.





Figure 5. Distance and Spot Size



## Field of View

Make sure that the target is larger than the unit's spot size. The smaller the target, the closer you should be to it. See Figure 6.



ame004.eps

Figure 6. Field of View

## Emissivity

Emissivity describes the energy-emitting characteristics of materials. Most organic materials and painted or oxidized surfaces have an emissivity of 0.95.

The 63 has a preset emissivity of 0.95. To compensate for inaccurate readings that may result from measuring shiny metal surfaces, cover the surface to be measured with masking tape or flat black paint (<148 °C/300 °F). Allow time for the tape or paint to

reach the same temperature as the surface beneath it. Measure the temperature of the tape or painted surface.

The 66/68 thermometers allow you to adjust the unit's emissivity for the type of surface being measured. Refer to Table 2.

**Table 2. Surface Emissivity**

| Measured Surface              | Emissivity | Measured Surface     | Emissivity |
|-------------------------------|------------|----------------------|------------|
| <b>METALS</b>                 |            | <b>Iron</b>          |            |
| <b>Aluminum</b>               |            | Oxidized             | 0.5-0.9    |
| Oxidized                      | 0.2-0.4    | Rusted               | 0.5-0.7    |
| <b>Alloy A3003</b>            |            | <b>Iron, Cast</b>    |            |
| Oxidized                      | 0.3        | Oxidized             | 0.6-0.95   |
| Roughened                     | 0.1-0.3    | Unoxidized           | 0.2        |
| <b>Brass</b>                  |            | Molten               | 0.2-0.3    |
| Burnished                     | 0.3        | <b>Iron, Wrought</b> |            |
| Oxidized                      | 0.5        | Dull                 | 0.9        |
| <b>Copper</b>                 |            | <b>Lead</b>          |            |
| Oxidized                      | 0.4-0.8    | Rough                | 0.4        |
| Electrical Terminal<br>Blocks | 0.6        | Oxidized             | 0.2-0.6    |
| <b>Haynes</b>                 |            | <b>Molybdenum</b>    |            |
| Alloy                         | 0.3-0.8    | Oxidized             | 0.2-0.6    |
| <b>Inconel</b>                |            | <b>Nickel</b>        |            |
| Oxidized                      | 0.7-0.95   | Oxidized             | 0.2-0.5    |
| Sandblasted                   | 0.3-0.6    | <b>Platinum</b>      |            |
| Electropolished               | 0.15       | Black                | 0.9        |

Table 2. Surface Emissivity (cont.)

| Measured Surface   | Emissivity | Measured Surface      | Emissivity |
|--------------------|------------|-----------------------|------------|
| <b>Steel</b>       |            | <b>Clay</b>           | 0.95       |
| Cold-Rolled        | 0.7-0.9    | <b>Concrete</b>       | 0.95       |
| Ground Sheet       | 0.4-0.6    | <b>Cloth</b>          | 0.95       |
| Polished Sheet     | 0.1        | <b>Glass</b>          |            |
| <b>Zinc</b>        |            | Plate                 | 0.85       |
| Oxidized           | 0.1        | <b>Gravel</b>         | 0.95       |
| <b>NON-METALS</b>  |            | <b>Gypsum</b>         | 0.8-0.95   |
| <b>Asbestos</b>    | 0.95       | <b>Ice</b>            | 0.98       |
| <b>Asphalt</b>     | 0.95       | <b>Limestone</b>      | 0.98       |
| <b>Basalt</b>      | 0.7        | <b>Paper</b>          |            |
| <b>Carbon</b>      |            | (any color)           | 0.95       |
| Unoxidized         | 0.8-0.9    | <b>Plastic</b>        |            |
| <b>Graphite</b>    | 0.7-0.8    | Opaque                | 0.95       |
| <b>Carborundum</b> | 0.9        | <b>Soil</b>           | 0.9-0.98   |
| <b>Ceramic</b>     | 0.95       | <b>Water</b>          | 0.93       |
|                    |            | <b>Wood (natural)</b> | 0.9-0.95   |

### Switching °C and °F

Open the unit by pushing the button on the underside of the thermometer near the trigger and pulling down and forward at the top of the handle.

To toggle between °C and °F, slide the top switch to the appropriate position. See Figure 2.


### **Trigger Lock (66/68)**



To lock the unit on for continuous measurement, slide the bottom switch down. See Figure 2. If the trigger is pulled while the unit is locked on, the laser and backlight will turn on if they have been activated. When the unit is locked on, the laser will turn off when the trigger is released. However, the backlight remains on unless it is turned off using the keypad.


### **Backlight and Laser On/Off Switch (63)**

To turn the backlight and laser on or off, slide the backlight and laser switch to the appropriate position. See Figure 2.

### **Function Button Functions (66/68)**

Models 66 and 68 measure Maximum (MAX), Minimum (MIN), Differential (DIF), and Average (AVG) temperatures each time a reading is taken. DIF shows the difference between maximum and minimum temperatures measured. AVG shows the average temperature reading for each time the trigger is pulled or the unit is locked on. This data is stored and can be recalled with  until a new measurement is taken. See “Hold” for information on how to recall stored data. When the trigger is pulled again, the unit will be measuring in the last mode selected.

Pressing  also allows access to the High Alarm (HAL), Low Alarm (LAL), Emissivity (EMS), Probe temperature (PRB—only available when the probe is connected to the thermometer), and Data logger (LOG). Each time  is pressed the unit advances the function cycle. Figure 7 shows the sequence of functions in the function cycle.



ame006.eps

**Figure 7. Function Cycle**

### **Selecting a Function (66/68)**

To select the MAX, MIN, DIF, or AVG mode, pull the trigger. While holding the trigger, press **FUNC** until the appropriate icon appears in the lower left corner of the display. Each time **FUNC** is pressed, the thermometer advances through the function cycle. The function cycle is shown in Figure 7.

## Setting the High Alarm, Low Alarm, and Emissivity (66/68)

To set values for the High Alarm (HAL), Low Alarm (LAL), and Emissivity, pull the trigger or press **FUNC** to activate the display. Press **FUNC** until the appropriate icon appears in the lower left of the display. Use **▲** or **▼** to adjust the desired values. To activate the alarms, press **ALARM**. To deactivate the alarms, press **ALARM** again.

## Using a Contact Temperature Probe (PRB) (66/68)

### ⚠ ⚠ Warning

**To avoid electrical shock or personal injury, do not connect the optional external probe to live electrical circuits.**

Connect the probe to the input on the side of the unit. PRB appears in the left of the display. The probe temperature appears in the lower right of the display. The live infrared temperature continues to show in the center of the display. While the probe is connected, you may still cycle through the mode functions by pressing **FUNC**. See Figure 8.

#### Note

*PRB is only available in the Function loop when a probe is connected to the unit. The probe temperature will not activate the high alarm or low alarm.*






Figure 8. Connecting a Temperature Probe





ame009.eps

## **HOLD**




The unit's display will remain activated for seven seconds after the trigger is released, unless the unit is locked on (lock on 66/68 only). HOLD appears in the upper left of the display. During HOLD or after the unit shuts off, recall stored values by pressing  without pulling the trigger (66/68). Each time  is pressed, the unit advances through the function cycle. When the trigger is pulled again, the unit will begin measuring in the last function selected.

## **Storing Data (66/68)**

The thermometers are capable of storing up to 12 data locations. The infrared temperature, temperature scale ( $^{\circ}\text{C}$  or  $^{\circ}\text{F}$ ), and emissivity are also stored.

To store data from an infrared reading, pull the trigger. While holding the trigger, press  until LOG appears in the lower left of the display. A log location number is shown below LOG. If no temperature has been recorded in the shown LOG location, 3 dashes will appear in the lower right corner. Aim the unit at the desired target and press . A tone confirms that the location temperature has been recorded. The recorded temperature appears in the lower right corner of the display. To select another log location, press  or .

## **Recalling Data (66/68)**

Recall stored data after the unit shuts off by pressing  until LOG appears in the display's lower left corner. A LOG location is shown below LOG and the stored temperature for that location will be shown in the display's lower right corner. To move to another LOG location, press  or .


## **LOG Clear Function**

The LOG Clear function allows you to quickly clear all logged data points. This function can only be used when the unit is in LOG


mode. It can be used when the thermometer has any number of log locations stored.


*Note*

*Only use the LOG Clear function if you want to clear all the LOG location data that is stored in the unit's memory.*

To use LOG Clear, while in LOG mode, pull the trigger then press  until the unit reaches LOG location "0".

*Note*

*This can only be done when the trigger is pulled. LOG location "0" cannot be accessed by using .*

When LOG location "0" shows in the display's lower left corner, press . Three tones sound and the LOG location automatically changes to "1", signifying that all data locations have been cleared.

## **Maintenance**

### **Changing the Battery**

To install or change the 9 V battery, open the unit and attach the battery to the battery snaps with the positive side facing toward the rear of the battery compartment. See Figure 2.

### **Cleaning the Lens**

Blow off loose particles using clean compressed air. Carefully wipe the surface with a moist cotton swab. The swab may be moistened with water.

### **Cleaning the Housing**


Use soap and water on a damp sponge or soft cloth.

** Caution**

**To avoid damaging the thermometers, do NOT submerge them in water.**



## Troubleshooting

| <b>Symptom</b>  | <b>Problem</b>  | <b>Action</b>   |
|---|---|---|
| --- (on display)  | Target temperature is over or under range   | Select target within specifications   |
|  | Low battery   | Replace battery   |
| Blank display   | Possible dead battery   | Check and/or replace battery  |
| Laser does not work   | <ol style="list-style-type: none"> <li>1. Low or dead battery</li> <li>2. Ambient temperature above 40 °C (104 °F)</li> </ol> | <ol style="list-style-type: none"> <li>1. Replace battery</li> <li>2. Use in area with lower ambient temperature</li> </ol> |
| ERR   | Possible damage by EMF  | Contact your distributor  |

## Accessories

Optional accessories for the thermometers are:

- Contact probe (RTD) (66/68) - Fluke PN 2148313
- Nylon Soft Carrying Case – Fluke PN 2152040
- NIST/DKD certification

## CE Certification

The thermometers conform to the following standards:

- EN61326-1 EMC
- EN61010-1
- EN60825-1 Safety

Certification testing was conducted using a frequency range of 80-1000 MHz with the instrument in three orientations.

### Notes

63: *Between 165 MHz and 880 MHz ( $\pm 5\%$ ) at 3 V/m, the instrument may not meet its stated accuracy.*

66/68: *Between 162 MHz and 792 MHz ( $\pm 5\%$ ) at 3 V/m, the instrument may not meet its stated accuracy.*

## Specifications

### Temperature Range

Assumes ambient operating temperature of 23 °C (73 °F) to 25 °C (77 °F)

|     |                                      |
|-----|--------------------------------------|
| 63: | -32 °C to 535 °C (-25 °F to 999 °F)  |
| 66: | -32 °C to 600 °C (-25 °F to 1100 °F) |
| 68: | -32 °C to 760 °C (-25 °F to 1400 °F) |

### Accuracy

Target Temperature:

|                      |  |
|----------------------|--|
| Above 510 °C (63)    | $\pm 1.5$ % of reading   |
| Above 510 °C (66/68) | $\pm 1$ % of reading or $\pm 1$ °C ( $\pm 2$ °F), whichever is greater |
| 23 °C to 510 °C      | $\pm 1\%$ of reading or $\pm 1$ °C ( $\pm 2$ °F), whichever is greater |
| -18 °C to 23 °C      | $\pm 2$ °C ( $\pm 3$ °F)   |
| -26 °C to -18 °C     | $\pm 2.5$ °C ( $\pm 4$ °F)   |
| -32 °C to -26 °C     | $\pm 3$ °C ( $\pm 5$ °F)   |

### Resolution

|        |                 |
|--------|-----------------|
| 63:    | 0.2 °C (0.5 °F) |
| 66/68: | 0.1 °C (0.1 °F) |

**Distance to Spot Ratio**

|     |      |
|-----|------|
| 63: | 12:1 |
| 66: | 30:1 |
| 68: | 50:1 |

**Emissivity**

(Adjustment for surface measured)

|        |                                       |
|--------|---------------------------------------|
| 63:    | Pre-set at 0.95                       |
| 66/68: | Digitally adjustable from 0.10 to 1.0 |

**Response Time**

500 mSec

**Repeatability**

±0.5 % of reading or ±1 °C (±2 °F), whichever is greater

**Spectral Response**

8 μm to 14 μm

**Laser Sighting**

Laser turns off above ambient temperature of 40 °C (104 °F)

**Relative Humidity**

10 % to 90 % RH non-condensing, at <30 °C (86 °F)

**Tripod Mount**

¼ in. 20 UNC threading

**Operating Temperature**

0 °C to 50 °C (32 °F to 120 °F)

**Storage Temperature**

-20 °C to 60 °C (-4 °F to 140 °F)

**Weight**

320 g (0.7 lbs)

**Dimensions**

200 mm x 160 mm x 55 mm (7.8 in x 6.30 in x 2.17 in)

**Battery Type**

9 V Alkaline or NiCd

**Battery Life**

|        |   |
|--------|---|
| 63:    | 10 hours with laser and backlight on<br>40 hours with laser and backlight off |
| 66/68: | 20 hours with laser and backlight on<br>40 hours with laser and backlight off |

## 63/66/68

### *Infrared Thermometers*

---

#### **Optional Contact Probe (66/68)**

**Temperature Range** -40 °C to 260 °C (-40 °F to 500 °F)

#### **Accuracy**

(Assumes ambient operating temperature of 23 °C (73 °F) to 25 °C (77 °F))

±1% of reading or ±1°C (±2°F), whichever is greater

**Certifications:** CE,



沪制01120009号